	A B Al	A
·	Application No.	Applicant(s)
Notice of Allowability	10/024,839	HEATH ET AL.
	Examiner	Art Unit
	Brian R. Gordon	1743
The MAILING DATE of this communication appears on the cover sheet with the correspondence address All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS. This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.		
1. This communication is responsive to <u>8-10-05</u> .		
2. The allowed claim(s) is/are <u>2-20</u> .		
 3. Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f). a) All b) Some* c) None of the: 1. Certified copies of the priority documents have been received. 2. Certified copies of the priority documents have been received in Application No 		
3. Copies of the certified copies of the priority documents have been received in this national stage application from the		
International Bureau (PCT Rule 17.2(a)).		
* Certified copies not received:		
Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application. THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.		
4. A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.		
5. CORRECTED DRAWINGS (as "replacement sheets") must be submitted.		
(a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached		
1) hereto or 2) to Paper No./Mail Date		
(b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date		
Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).		
6. DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.		
Attachment(s)	_	
1. Notice of References Cited (PTO-892)		atent Application (PTO-152)
2. Notice of Draftperson's Patent Drawing Review (PTO-948)	 Interview Summary Paper No./Mail Date 	
3. Information Disclosure Statements (PTO-1449 or PTO/SB/08 Paper No./Mail Date	8), 7. Examiner's Amendm	
4. Examiner's Comment Regarding Requirement for Deposit	8. 🛛 Examiner's Stateme	nt of Reasons for Allowance
of Biological Material	9.	

DETAILED ACTION

Allowable Subject Matter

- 1. Claims 2-20 are allowed.
- 2. The following is an examiner's statement of reasons for allowance: The prior art of record does not teach nor fairly suggest: a mixing and pouring apparatus, comprising: a base; a locking arm support carried on the base; a locking arm rotatably mounted within the locking arm support; and a drive mechanism operatively coupled to the locking arm, the drive mechanism capable of rotating the locking arm drive mechanism comprises a motor having a drive shaft, the motor connectable to an external motor control; a drive gear operatively coupled to the drive shaft; a free gear operatively. fixedly coupled to the rotatable locking arm; and a belt seated over the drive gear and the free gear, and wherein the belt is movable to drive the free gear in response to motion of the drive gear; a registration mechanism, the registration mechanism comprising: a registration disk operatively, fixedly coupled to the free gear, the registration disk having a registration slot therein; an optocoupler having a transmitter and a receiver separated by a gap, wherein the registration disk is positioned to extend into the gap; and control lines operatively electrically connected to the optocoupler and to the motor; and wherein the registration slot is aligned in the gap of the optocoupler when the registration disk is in a home position wherein the locking and is in a substantially vertical position.

The prior art further fails to teach or suggest a locking arm further comprising a plurality of vessel openings and a matching plurality of vacuum ports, each of the vessel

Art Unit: 1743

openings sized to accommodate a vessel, and each of the locking ports capable of retaining the vessel in the locking arm; a drain trough for receiving waste material from a vessel situated in the locking arm when the locking arm is rotated to pour material from a vessel; and wherein the base includes a plurality of guide pin openings, the apparatus further comprising: a supplemental cradle having a plurality of cradle vessel openings each sized to accommodate a vessel, the supplemental cradle having a plurality of guide pins extending therefore to engage the guide pins with the guide pin openings to position the supplemental cradle on the base.

3. Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Brian R. Gordon whose telephone number is 571-272-1258. The examiner can normally be reached on M-F, with 2nd and 4th F off.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Jill Warden can be reached on 571-272-1267. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Application/Control Number: 10/024,839

Art Unit: 1743

Page 4

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

brg

// Jill Warden
Supervisory Patent Examiner
Technology Center 1700